

Data Sheet

RISH Max-14 Analog-Digital Multimeter





RISHABH INSTRUMENTS PVT.LTD. F-31, MIDC, Satpur, Nashik-422 007,India. Tel.: +91 253 2202160, 2202202 Fax: +91 253 2351064 E-mail : India :- marketing@rishabh.co.in International :- exp.marketing@rishabh.co.in www.rishabh.co.in

Rishabh Instruments

- Input resistance can be selected for voltage measurements.
- Direct and alternating voltages from 400mV ... 600V.
- Direct and alternating currents from 40mA ... 10A.
- Resistance from 400Ω ... 40MΩ.
- Capacitance from 4nF ... 40µF with relative operation.
- · Frequencies from 10Hz to 400KHz.
- Diode measurement and continuity testing.
- MIN. MAX and Hold measurement storage.

Application

RISH max digital multimeters are suited for universal, general applications in the electrical and electronics fields, as well as in radio and television service, training and education.

They are of especially flat design, and thus fit into any bag. The protective cover, which is provided as standard accessory, can be opened at an angle for convenient reading from the workbench, and provides for easy transport.

Selection of input resistance for voltage measurement

In addition to the usual voltage input with one resistance value of 10 M Ω , which is selected via ~ or V=-, this measuring instrument provides the electrician with an additional selector switch position for $V_{400K\Omega}$ with an input resistance of approx. 400 k Ω . This allows for the avoidance of negative influences from capacitive coupling during voltage measurements in power supply systems.

Effective value for distorted waveform

The built-in effective value transducer allows for effective value measurement (TRMS) indepedent of waveform for alternating magnitudes (AC).

Hold

By pressing the HOLD/ON key, the currently displayed measurement value can be held and "HOLD" is simultaneously displayed.

Min/Max

The minimum and maximum values which were present at the input of the measuring instrument after activation of the MIN MAX mode can be selectively 'retained' with the MIN MAX function. The most important application is the determination of the minimum or maximum value during long term observation of measurement quantities. MIN/MAX has no effect on the analog display; it continues to display the current measurement value.

Automatic/manual measuring range selection

The measurement quantities are chosen with the rotary selector switch. The measuring range is automatically adjusted to the measurement value. The measuring range can also be manually selected with the AUTO/MAN button.

· Diode and continuity testing

This provides for the testing of the polarity of diodes, as well as inspection for short-circuits and circuit interruptions. In addition to the display, resistance of less than 40 Ω are indicated with an acoustic signal.

Overload warning

An acoustic signal occurs, if the range limit value is exceeded.

Energy saving circuit

The instrument is switched off automatically, if none of the operating elements have been activated for about 30 minutes.

• Protective cover for rough operating conditions

A protective cover of ABS with a built-in stand protects the instrument against jolts and falls. It also secures the test probes for one-hand operation, and allows for winding of the measurement cable which provides protection during transport.

Calibration

Rish max series multimeters are calibrated using Fluke 5500 & Wavetek 9100. These sources are calibrated at regular intervals.

Theft protection

Company name and name of the user can be entered into the field next to the display with an indelible etching needle for identification of the owner.

RISHABH INSTRUMENTS PVT.LTD. INSTRUMENTS RISHABH Measure, Control & Record with a Difference www.rishabh.co.in

Page 2 of 5

F-31, MIDC, Satpur, Nashik-422 007,India. Tel.: +91 253 2202160, 2202202 Fax : +91 253 2351064 E-mail : India :- marketing@rishabh.co.in International :- exp.marketing@rishabh.co.in

Rishabh Instruments

1 / 05 / IX / 1210 / B

Characteristic values for Rish Max 14

Meas. Function	Measuring Range	Resolution	Input Impedance 100 pF//XΩ		Digital display inherent deviation at reference condition	Overload capacity ¹⁾	
			V == / ~	V 400ΚΩ	<u>+(%</u> of rdg. +digits)	Overload value	Overload Duration
	400.0mV	100μV	>20MΩ	~400KΩ	0.75+2		
V ==	4.000V	1mV	11MΩ	~400KΩ			
	40.00V	10mV	10MΩ	~400KΩ	0.5+2	720V 	Continuous
V ==-	400.0V	100mV	10MΩ	~400KΩ			
	600V	1V	10MΩ	~400KΩ			
	400.0mV ²⁾	100μV	>20MΩ	~400KΩ	1.5+5 ³⁾		
٧~	4.000V ²⁾	1mV	11MΩ	~400KΩ	3)		
	40.00V ²⁾	10mV	10MΩ	~400KΩ	1+5 ′	720V ~ effective sine	Continuous
V~ _{400КΩ}	400.0V ²⁾	100mV	10MΩ	~400KΩ		checuve sine	
	600V ²⁾	1V	10MΩ	~400KΩ	1+10 ³⁾		
			Approx. vo at max. me	oltage drop as. current			
	40.00mA	10µA	450	mV	0.010	400 4	
Δ	400.0mA	100µA	1.5	V	0.8+2	480MA	Continuous
	10.00A ⁶⁾	10mA	750	mV	1.5+5	6)	6)
	40.0mA ²⁾	10µA	450	mV	³⁾ 1+5	480mA	Continuous
A~	400.0mA ²⁾	100µA	1.5	V		1001181	
	10.00A ^{6) 2)}	10mA	750	mV	2+5 ³⁾	6)	6)
			Open - circ	cuit voltage			
	400.0Ω	100mΩ			0.8+5		
Ω	4.000ΚΩ	1Ω				420 V	
	40.00ΚΩ	10Ω	annrox	0.51/	0.8+2	DC/AC	
	400.0ΚΩ	100Ω	approva	0.01		effective	10 min
	4000ΚΩ	1ΚΩ			1+5	3110	
	40.00MΩ	10KΩ			2+5		
BUZZER	400.0Ω	100mΩ			Acoustic signal for 0< 40 Ω		
DIODE	3.000V	1mV	approx	. 3V	2+10		
	4.000nF	1pF			3+40 4)	420 \/	
	40.00nF	10pF			3+10 ⁴⁾	420 V DC/AC	10 min
F	400.0nF	100pF			3+10	effective	TO THIT
	4.000μF	1nF			5.10	sine	
	40.00μF	10nF			5+10		
			fmir	ı			
	100.00Hz	0.01Hz	10H	z		<u>≤</u> 1KHz : 600V	
	1.0000KHz	0.1Hz	10H	z			
ы ₅)	10.000KHz	1Hz	10H	z	0.2+2	<u>≤</u> 10KHz : 400V	Continuous
	100.00KHz	10Hz	10H	z	z		
	400.0KHz	100Hz	100	Hz		<u>- 1001012 . 400</u>	

2) Effective value measurement (TRMS) for *RISH Max* 14 TRMS measurement is independent of waveform.

3) The specified inherent deviation is valid for RISH Max 14 4) With zero adjustment 'REL'; without zero adjustment +300 digits in 40nF range
4) With zero adjustment 'REL'; without zero adjustment 'A00 digits in 40nF range

5) Indication of the frequency measurement expanded up to 9999 digits. 6) max. 10 A/30 min

RISHABH

Measure, Control & Record with a Difference

- 12 A/5 min
- 16 A/30 sec

Applicable regulations and standards

IEC 1010-1 DIN EN 61010 Part 1 VDE 0411 - 1	Safety regulations for electrical measuring, control, regulation and laboratory devices	
DIN 43751 IS 13875	Digital measuring instruments Digital measuring instruments	
DIN EN 50081 Part 1	Generic emmision standard residential, commercial and light industry	
DIN EN 50082 Part 1	Generic immunity standard residential, commercial and light industry	
VDI/VDE 3540	Reliability of measuring, control and regulation instruments	
DIN EN 60529 DIN VDE 0470 Part 1	Test instruments and test procedures Degree of protection provided by enclosures (IP code)	Page 3 of 5
		-



RISHABH INSTRUMENTS PVT.LTD. F-31, MIDC, Satpur, Nashik-422 007, India. Tel.: +91 253 2202160, 2202202 Fax: +91 253 2351064 E-mail : India :- marketing@rishabh.co.in International :- exp.marketing@rishabh.co.in



www.rishabh.co.in

Rishabh Instruments

Reference Conditions

Ambient temperature	+ 23 °C ± 2 K
Relative humidity	45 % 55 %
Frequency of	
meas. quantity	Sine 50 Hz
Operating voltage	$8 V \pm 0.1 V$

Display

LCD display field (50 mm x 30 mm) with analog and digital display and with display of measurement unit, type of current and various special functions.

Digital

Display	7 segment
Character height	10 mm
Number of digits	3 3/4 digit
Overflow display	,4000' with blinking ,4'
Polarity display	,–' sign is displayed when plus pole at ,⊥ '
Measurement rate	3 measurement/s for V, I, Ω .
	1 measurement/s for capacitive and
	frequency measurements.

Analog

Display	LCD scale with bar graph disp	olay		
Scale lengtht	40 mm			
Scaling Deterite diseter				
Polarity display	with automatic reversal			
Overflow display	Bar with triangle			
Measurement rate	20 measurement/s			
2	3	4	5	
4 1	1			



RISH Max display :

- 1 Digital display with comma and polarity display
- 2 Low Battery Indication
- 3 Display for REL and HOLD as well MIN MAX storage
- 4 Continuity test display:
- speaker symbol appears when acoustic signal is switched on 5 Display for diode measurement
- 6 Measurement unit display
- 7 Display for exceeding of measuring range
- 8 Indicator for analog display
- 9 Scale for analog display
- 10 Display for automatic or manual measuring range selection
- 11 Display for selected type of current (AC or DC)



Rishabh Instruments





For unknown waveform (crest factor CF > 2) measurement to be made with manual range selection
 Except for sine waveform

Influence variable	Influence range (max. resolution)	Frequency	Inherrent Error at Ref. (%rdg +digits)
Frequency	4, 40, 400V	20Hz <50Hz >50Hz 500Hz	2 + 3
Vac	400mV, 600V	20Hz <50Hz >50Hz 100Hz	2 + 3

Influence	Influence	Meas. Quantity /	Influence
Variable	Range	Meas. Range	Effect
Relative humidity	55 75%	$V \simeq$ A \simeq Ω F Hz	

Influence Variable	Interference Magnitude	Meas. Quantity / Meas. Range	Attenuation
	600V DC/AC 50Hz sinusoidal	All V DC	>100 dB
Common	600V DC	All V DC	>100 dB
Mode	600V AC 50Hz sinusoidal	400mV / 4V AC	>80 dB
Voltage		40V AC	>63 dB
-		400V AC	>43 dB
		600V AC	>23 dB
Series - Mode	AC 50/60 Hz	V DC	>43 dB
voltage	MAX. 600V DC	V AC	>55 dB

Aux. Voltage Influence (without ⊣⊢ display)

all ranges except AC : <u>+</u>5 D AC range : <u>+</u>20 D

Page 4 of 5

RISHABH INSTRUMENTS PVT.LTD. F-31, MIDC, Satpur, Nashik-422 007,India. Tel.: +91 253 2202160, 2202202 Fax : +91 253 2351064 E-mail : India :- marketing@rishabh.co.in International :- exp.marketing@rishabh.co.in

www.rishabh.co.in

1 / 05 / IX / 1210 / B

Power Supply

Battery	9 V flat cell battery Zinc-carbon cell per IEC 6 F 22 Alkaline manganese dry cell per IEC 6 LR 61
Service life	Zinc-carbon cell: approx. 150 hours Alkaline manganese dry cell approx. 300 hours
Battery test	Automatic display of " \perp " symbol when battery voltage falls below following value: approx. 7V

Electromagnetic Compatibility

Interference emission

Interference immunity 10 A range

EN 50081-1: 1992 EN 55022: 1987 class B EN 50082-1 : 1992 EN 61000-4-2:8kV air discharge EN 61000-4-3: 3 V/m EN 61000-4-4; 0.5 kV

Ambient Conditions

Operating
temperature range-10 °C ... + 50 °CStorage temperature range- 25 °C ... + 70 °C
(without batteries)Climate classification2z/-10/50/70/75%
in correspondence with
VDI/VDE 3540Relative humidity45 ... 75 %
up to 2000 m

Fusing

 $\begin{array}{ll} \mbox{Fwse for ranges} & \mbox{FF 1.6 / 500 V; 6.3 mm x 32mm} \\ \mbox{Breaking capacity 50 kA at 500 V} \sim \mbox{and} \\ \mbox{non-reactive load, } \cos \phi < 0.2; \mbox{ protects all} \\ \mbox{current measuring ranges up to 400 mA in} \\ \mbox{connection with power diodes} \end{array}$

Fuse for FF 16 A / 500 V; 6.3 mm x 32 mm 10 A range breaking capacity 50 kA at 500 V ~ and non-reactive load, $\cos \varphi < 0.2$

Mechanical Design

Protection	Instruments: IP 50 Connector sockets: IP 20
Dimensions	W x H x D: 92 mm x 154 mm x 25 mm
Weight	Approx. 0.2 Kg with battery

Electrical safety

Protection classII per IEC 1010-1/EN 61010-1/VDE 0411-1OvervoltageIIClassificationIINominal voltage600 VContamination level222Test voltage3.7 kV ~IEC 1010-1/EN 61010-1VDE 0411-1

Included equipment

- 1 Multimeter
- 1 Probe set

1 Copy Operating Instructions

1 Protective Case with tilt stand

Designation	Order Code
Analog-Digital multimeter with TRMS	33050
RISH Max Fuse 1.6A	42124
RISH Max Fuse 16A	42198
RISH Max Probe Set	42199
Safety cover RISHmax 14	42200





RISHABH INSTRUMENTS PVT.LTD.

F-31, MIDC, Satpur, Nashik-422 007,India. Tel.: +91 253 2202160, 2202202 Fax : +91 253 2351064 E-mail : India :- marketing@rishabh.co.in International :- exp.marketing@rishabh.co.in

www.rishabh.co.in

Page 5 of 5

1 / 05 / IX / 1210 / B