

RISH Multi 20 CAT IV

5 3/4 Digits High Precision Digital Multimeter

Product Introduction:

- 300,000 digits and triple display
- Precision multimeter (V, A, Ω , Hz, F, °C, Diode)
- Direct current measurement (For 10A)
- Safety: 1000V CAT III/600V CAT IV

Applications :

RISH Multi 20 Multimeter is 5 3/4 digit high performance precision instruments for laboratory use, as well as for service and training . With a display range of 300,000 digits, as well as outstanding accuracy and long-term stability, they fulfill all requirements for use in calibration and R&D laboratories. The instruments can be utilized on-site for precision maintenance and calibration tasks

Features :

Convenient Triple Display

Simultaneous measurement of voltage and frequency or voltage and min, max values is possible. Also current and frequency or the current measurement and minimum and maximum values are displayed simultaneously. The usual switching back and forth between display values is no longer necessary. Main display resolution for AC is 4 3/4 places and for DC it is 5 3/4 places

High Resolution & Precision

5 3/4 places (309,999 digits) for DC measured quantities and 4 3/4 places (30,999 digits) for AC measured quantities allow for precision reference measurements and use as a calibration standard for testing devices and assemblies.

RMS Value with Distorted Waveshape

The utilized measuring method allows for waveshape independent RMS measurement (TRMSAC)

Additional Functions

Resistance (Ω), Temperature (°C), Continuity, Diode, Capacitance, Frequency, Duty cycle

Overload Protection

The instrument is safeguarded for up to 1000 V in all measuring functions by overload protection. An acoustic signal is generated if the upper voltage or current range limit is exceeded. FUSE appears at the display if the fuse for the current

Measuring Category

1000V CAT III/600V CAT IV as per IEC Standard 61010-1.

RISH Multi 20 CAT IV



Automatic Blocking Sockets (ABS)

Automatic blocking sockets prevent incorrect connection of measurement cables and inadvertent selection of the wrong measured quantity. This significantly reduces danger to the user, the instrument and the system under test, and in many cases eliminates it entirely.

Backlit Display

The instrument is provided with user selectable Back-lit for taking measurements in poor lighting conditions / dark areas.

Applicable Regulations and Standards :

IEC 1010-1 /EN61010-1	Safety regulations for electrical measurement, Control, regulating and laboratory devices
IEC 61326:2002 Class B	Electrical equipment for control technology and laboratory use – EMC requirements
IEC 61000-4-2	8kV atmosphere discharge. 4kV contact discharge
IEC 6100-4-3	3VT/m
DIN EN 60259 DIN VDE 0470 part 1	Test equipment and test procedures -Degrees of protection provided by enclosures (IP Code).

Subject to change without notice



Rishabh Instruments

RISHABH
INSTRUMENTS
Measure, Control & Record with a Difference

RISHABH INSTRUMENTS PVT.LTD.

F-31, MIDC, Satpur, Nashik-422 007, India.

Tel.: +91 253 2202160, 2202202 Fax : +91 253 2351064

E-mail : India :- marketing@rishabh.co.in

International :- exp.marketing@rishabh.co.in

www.rishabh.co.in

RISH Multi 20 CAT IV

5 3/4 Digits High Precision Digital Multimeter

Reference conditions for Accuracy:

Reference temperature	23°C ± 2K
Relative Humidity	45%...55% RH
Waveform of measured quantity	Sinusoidal
Input frequency	50 or 60 Hz ±2%
Battery Voltage	8 V ± 0.1 V

Applicable Standards:

EMC	IEC 61326: Class B
Immunity	IEC 61000-4-2 8 KV atmosphere discharge, 4 KV contact discharge IEC 61000-4-3 : 3 V/m
IP for water & dust	IEC60529 : IP 50 for Housing

Safety

	IEC 61010-1-2001
Pollution degree:	2
Installation category:	III IV
Operating voltage	1000V 600V
High Voltage Test	6.7 kV (IEC 61010-1-2001)

Environmental

Operating temperature	-10 to +50°C
Storage temperature	-25 to +70°C
Relative humidity	<75% non condensing.
Altitude	Up to 2000 m

Mechanical Configuration

Protection type	IP 50, for the connection sockets IP 20.
Dimensions	84 mm X 195 mm X 35mm
Weight	350 gm approx, including battery.

Fuses

For ranges up to 300 mA	FF1.6A /1000V/6.3mm x 32 mm. Rating 20 KA with 1000 V~ and ohmic load. In conjunction with diodes protects all current measuring ranges up to 300 mA.
For the range 10A	16 A /1000V /10 mm X 38 mm, protects 10 A range up to 1000 V.

Battery

Battery Voltage	9V flat cell battery
Battery type	Manganese Dioxide cell as per IEC 6 F22 or alkaline-manganese cell according to IEC 6 LR61 suitable NiCd storage battery.
Battery Life	Minimum 300 hours on Vdc, Adc ,150 hours on Vac, Aac. Without Backlit

Subject to change without notice



Rishabh Instruments

RISHABH
INSTRUMENTS
Measure, Control & Record with a Difference

RISHABH INSTRUMENTS PVT.LTD.

F-31, MIDC, Satpur, Nashik-422 007, India.

Tel.: +91 253 2202160, 2202202 Fax : +91 253 2351064

E-mail : India :- marketing@rishabh.co.in

International :- exp.marketing@rishabh.co.in

www.rishabh.co.in

RISH Multi 20 CAT IV

5 3/4 Digits High Precision Digital Multimeter

Specifications :

Meas. function	Meas. Range	Resolution at Measuring Range Upper Limit		Input impedance		Inherent Deviation at highest resolution under reference condition		Overload Capacity 4)		Meas. function
				DC	AC	DC Accuracy ±(..% of rdg + .. %rng +...Digit)	AC Accuracy ⁹⁾ ±(..% of rdg +...Digit)	Overload Value	Overload Duration	
V	3.0V	100 µV	100 µV	11 MΩ	5 MΩ // < 50pF	0.02+0.008+20	0.2+30	DC AC RMS Sine 1000V	Continuous	V
	30.0V	1 mV	1 mV	10 MΩ	5 MΩ // < 50pF	0.02+0.008+20	0.2+30			
	300V	10mV	10 mV	10 MΩ	5 MΩ // < 50pF	0.02+0.008+20	0.2+30			
	1000V	100 mV	100 mV	10 MΩ	5 MΩ // < 50pF	0.02+0.008+20	0.2+30			
	300mV	10 µV	10 µV	>20MΩ	5 MΩ // < 50pF	0.02+0.015+30 6)	0.5+30			
A				Voltage drop. approx. for upper range limit B		DC Accuracy ±(..% of rdg + .. %rng +...Digit)	AC Accuracy ⁹⁾ ±(..% of rdg +...Digit)	0.36A	Continuous	A
	300µA	10nA	10nA	300mV	300mV	0.05+0.02+20	0.5+30			
	3 mA	100nA	100nA	300mV	300mV	0.02+0.01+20	0.5+30			
	30mA	1nA	1 µA	300mV	300mV	0.02+0.01+20	0.5+30			
	300mA	10 µA	10 µA	300mV	300mV	0.1+0.01+20	0.5+30			
10A	1mA	1mA	400mV	400mV	0.2+0.05+30	0.5+30	10A5)	5 min		
Ω				Open-circuit voltage		±(..% of rdg + .. %rng +...Digit)		1000V DC AC RMS	10 Sec	Ω
	300Ω	10mΩ		1.3V	--	0.05+0.015+20 6)				
	3.0 kΩ	100mΩ		0.5V	--	0.05+0.015+20				
	30 kΩ	1Ω		0.5V	--	0.05+0.015+20				
	300 kΩ	10Ω		0.5V	--	0.05+0.025+20				
	3.0 MΩ	100Ω		0.5V	--	0.1+0.025+20				
)	300 Ω	0.1 Ω		max. 1.3V	--	1+0.25+20		Sine)
	→	2.0 V	100µV	max. 2.5 V		0.2+0+10				
F				Discharge resistor	U ₀ max.	±(..% of rdg +... %rng)		1000V DC AC RMS Sine	10 Sec	F
	3.0 nF	1 pF		10 MΩ	3 V	2.5+0.2 ⁸⁾				
	30 nF	10pF		10 MΩ	3 V	1.2+0.2				
	300nF	100pF		1 MΩ	3 V	1.2+0.2				
	3.0µF	1 nF		100 kΩ	3 V	1.2+0.2				
	30µF	10nF		11 kΩ	3 V	1.2+0.2				
	300µF	100nF		2 kΩ	3 V	3.2+1				
	3000µF	1 µF		2 kΩ	3 V	3.2+1				
30000µF	10 µF		2 kΩ	3 V	3.2+1					
Hz				f min 3)		±(..% of rdg. +...Digit)		≤ 3 KHz 1000V ≤ 30 KHz 300V ≤ 100 KHz 30V	Continuous	Hz
	300.00Hz	0.01Hz		10 Hz		0.1+3 ⁷⁾				
	3.0000kHz	0.1Hz		100 Hz		0.1+3 ⁷⁾				
	30.0000kHz	1Hz		100Hz		0.1+3 ⁷⁾				
)	300.000kHz	10Hz		1kHz		0.05+10 ⁷⁾		1000V)
	→	100 min ²⁾	10 ms			±20Digit				
°C						±(..% of rdg. +...Digit)		1000V DC AC RMS Sine	10 Sec	°C
	Pt100	-200.0... +100.0 °C	0.1 °C			1Kelvin +3 ⁸⁾				
		+100.0... +850.0 °C				1% + 3 ⁸⁾				
	Pt1000	-100.0... +100.0 °C	1 Kelvin + 3 ⁸⁾							
+100.0... +850.0 °C		1% + 3 ⁸⁾								

- 1) Display : 5 3/4 places for DC , 4 3/4 places for AC
- 2) Stopwatch : Format : mm : ss : hh
Where m = minute, s = second and h = hundredth second, max. : 99 : 59 : 59
- 3) Lowest measurable frequency with sinusoidal measurement signal symmetric to zero point.
- 4) at 0 40 °C
- 5) 12A – 5min, 16A – 30 s
- 6) With zero adjuster; Without zero adjuster.
- 7) Vac >1V_{eff/rms}
- 8) Without sensor
- 9) Values less than 150 counts are suppressed. .



RISHABH INSTRUMENTS
Measure, Control & Record with a Difference

RISHABH INSTRUMENTS PVT.LTD.
F-31, MIDC, Satpur, Nashik-422 007, India.
Tel.: +91 253 2202160, 2202202 Fax : +91 253 2351064
E-mail : India :- marketing@rishabh.co.in
International :- exp.marketing@rishabh.co.in
www.rishabh.co.in

RISH Multi 20 CAT IV

5 3/4 Digits High Precision Digital Multimeter

Influence Quantities & Influence Errors:

Influence Quantity	Range of Influence	Measured Quantity/ Measuring Range	Variation ¹⁾ ± (...% of rdg. +digits)
Temperature	-10 °C +21 °C and +25 °C...+40°C	VDC	0.05+3
		V ~	0.2+3
		300µA ... 300mA DC	0.2+3
		300µA ... 300mA AC	0.3+3
		10 A AC, 10 A DC	0.5+3
		300 Ω ²⁾	0.1+5
		3 KΩ – 3 MΩ	0.1+3
		30 MΩ	0.6+3
		30 nF ²⁾ – 30 µF	0.5+3
		30 µF-30mF	2.0+3
		Hz	0.1+3
		-200...+200 °C	0.5 Kelvin+2 Digits
		+200...+850°C	0.5+2
		25 Hz...< 45 Hz	300 mV ~
> 65 Hz... 200 Hz		1.0+20	
Frequency of the measured quantity	25 Hz...< 45 Hz	3...300V ~	1.0+20
	> 65 Hz... 400 Hz		0.5+20
	>400 Hz...1 <KHz		0.5+20
	>1 KHz ...20 Khz		0.5+20
	25 Hz...< 30 Hz		1.0+20
	> 30 Hz... 45 Hz	1000V ~	0.5+20
	>65 Hz...1 <Khz		2.0+20
	25Hz ...< 45 Hz		1.0+20
	>65 Hz... 1 kHz	A~	1.0+20
Wave form of the measured quantity ³⁾	Crest factor CF	1...3	± 1 % of rdg
		3...5	± 3 % of rdg
Battery Voltage	+ ⁵⁾ ...< 7.5 V > 8.1 V ...10.0 V	V DC	±15 Digit
		V~	±30 Digit
		ADC	±20 Digit
		A AC	±40 Digit
		30Ω / 300 Ω/°C	±40 Digit
		3 kΩ – 30MΩ	±40 Digit
		Cpacitance	±50 Digit
		Hz	±10 Digit
Relative humidity	75% 3Days Meter off		±1 Digit
DATA	-		± 1 digits
MIN/MAX	-	V ac/dc , A ac/dc	± 2 digits



RISHABH INSTRUMENTS
Measure, Control & Record with a Difference

RISHABH INSTRUMENTS PVT.LTD.

F-31, MIDC, Satpur, Nashik-422 007, India.

Tel.: +91 253 2202160, 2202202 Fax : +91 253 2351064

E-mail : India :- marketing@rishabh.co.in

International :- exp.marketing@rishabh.co.in

www.rishabh.co.in

RISH Multi 20 CAT IV

5 3/4 Digits High Precision Digital Multimeter

Influence Quantity	Range of Influence	Measuring Ranges	Attenuation
Common Mode interference Voltage	Noise quantity max. 1000 V	V dc	> 120 dB
	Noise quantity max. 1000 V ~ 50 Hz, 60 Hz sinusoidal	300 mV~ ... 30V~	> 80 dB
		300 V~	> 70 dB
Normal Mode interference voltage	Noise quantity V ~ Value of the measuring range at a time Max. 1000V~ ,50Hz, 60Hz Sinusoidal	1000 V~	> 60 dB
		V dc	> 50dB
	Noise quantity max. 1000 V-	V~	>110dB

Response time (After manual range selection):

Measured Quantity/ Measured range	Response time	Transient response for step function of the measured quantity
VDC, VAC, AAC+DC, AAC	1.5 s	From 0 to 80 % of upper range limit.
30Ω...3 MΩ	2 s	
30 MΩ	5 s	From ∞ to 50 % of upper range limit.
→†	1.5s	
°C	3 s	From 0 to 50 % of upper range limit.
3.0nF,300μF,	Max. 1... 3 s	
3000 μF	Max. 7 s	
30000 μF	Max. 14 s	
300 Hz,3KHz	Max 2 s	
30 KHz,300 KHz	Max 0.7 s	



RISHABH
INSTRUMENTS
Measure, Control & Record with a Difference

RISHABH INSTRUMENTS PVT.LTD.

F-31, MIDC, Satpur, Nashik-422 007, India.

Tel.: +91 253 2202160, 2202202 Fax : +91 253 2351064

E-mail : India :- marketing@rishabh.co.in

International :- exp.marketing@rishabh.co.in

www.rishabh.co.in